

QNTSK Surgical Kit Reusable Drill Usage Chart

Use a dry erase marker to track the total number of osteotomy preparations that are performed with each drill. It is recommended that reusable drills be replaced after 15 uses*. Single use drills should be disposed of after each procedure.



| ACTPSD | ACT2010 | ACT2015 | ACT2020 | ICD100 | QSD3285 | QSD3210 | QSD3211 | QSD3213 | QSD3215 | QSD485 | QSD410 | QSD411 | QSD413 | QSD415 | QSD585 | QSD510 | QSD511 | QSD513 | QSD515 | QSD685 | QSD610 | QSD611 | QSD613 | QSD615 |
|-------------------------------------|---------------------------------------|---------------------------------------|---------------------------------------|------------------------------|------------------------------------|-----------------------------------|-------------------------------------|-----------------------------------|-----------------------------------|------------------------------------|-----------------------------------|-------------------------------------|-----------------------------------|-----------------------------------|------------------------------------|-----------------------------------|-------------------------------------|-----------------------------------|-----------------------------------|------------------------------------|-----------------------------------|-------------------------------------|-----------------------------------|-----------------------------------|
| ACT® Pointed Starter Drill | ACT Twist Drill 2 mm x 10 mm | ACT Twist Drill 2 mm x 15 mm | ACT Twist Drill 2 mm x 20 mm | Countersink Drill 4 mm | Quad Shaping Drill 8.5 mm | Quad Shaping Drill 10 mm | Quad Shaping Drill 11.5 mm | Quad Shaping Drill 13 mm | Quad Shaping Drill 15 mm | Quad Shaping Drill 8.5 mm | Quad Shaping Drill 10 mm | Quad Shaping Drill 11.5 mm | Quad Shaping Drill 13 mm | Quad Shaping Drill 15 mm | Quad Shaping Drill 8.5 mm | Quad Shaping Drill 10 mm | Quad Shaping Drill 11.5 mm | Quad Shaping Drill 13 mm | Quad Shaping Drill 15 mm | Quad Shaping Drill 8.5 mm | Quad Shaping Drill 10 mm | Quad Shaping Drill 11.5 mm | Quad Shaping Drill 13 mm | Quad Shaping Drill 15 mm |
| | | | | | 3.25 mmD | | | | | 4 mmD | | | | | 5 mmD | | | | | 6 mmD | | | | |
| 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1 |
| 2 | 2 | 2 | 2 | 2 | 2 | 2 | 2 | 2 | 2 | 2 | 2 | 2 | 2 | 2 | 2 | 2 | 2 | 2 | 2 | 2 | 2 | 2 | 2 | 2 |
| 3 | 3 | 3 | 3 | 3 | 3 | 3 | 3 | 3 | 3 | 3 | 3 | 3 | 3 | 3 | 3 | 3 | 3 | 3 | 3 | 3 | 3 | 3 | 3 | 3 |
| 4 | 4 | 4 | 4 | 4 | 4 | 4 | 4 | 4 | 4 | 4 | 4 | 4 | 4 | 4 | 4 | 4 | 4 | 4 | 4 | 4 | 4 | 4 | 4 | 4 |
| 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 |
| 6 | 6 | 6 | 6 | 6 | 6 | 6 | 6 | 6 | 6 | 6 | 6 | 6 | 6 | 6 | 6 | 6 | 6 | 6 | 6 | 6 | 6 | 6 | 6 | 6 |
| 7 | 7 | 7 | 7 | 7 | 7 | 7 | 7 | 7 | 7 | 7 | 7 | 7 | 7 | 7 | 7 | 7 | 7 | 7 | 7 | 7 | 7 | 7 | 7 | 7 |
| 8 | 8 | 8 | 8 | 8 | 8 | 8 | 8 | 8 | 8 | 8 | 8 | 8 | 8 | 8 | 8 | 8 | 8 | 8 | 8 | 8 | 8 | 8 | 8 | 8 |
| 9 | 9 | 9 | 9 | 9 | 9 | 9 | 9 | 9 | 9 | 9 | 9 | 9 | 9 | 9 | 9 | 9 | 9 | 9 | 9 | 9 | 9 | 9 | 9 | 9 |
| 10 | 10 | 10 | 10 | 10 | 10 | 10 | 10 | 10 | 10 | 10 | 10 | 10 | 10 | 10 | 10 | 10 | 10 | 10 | 10 | 10 | 10 | 10 | 10 | 10 |
| 11 | 11 | 11 | 11 | 11 | 11 | 11 | 11 | 11 | 11 | 11 | 11 | 11 | 11 | 11 | 11 | 11 | 11 | 11 | 11 | 11 | 11 | 11 | 11 | 11 |
| 12 | 12 | 12 | 12 | 12 | 12 | 12 | 12 | 12 | 12 | 12 | 12 | 12 | 12 | 12 | 12 | 12 | 12 | 12 | 12 | 12 | 12 | 12 | 12 | 12 |
| 13 | 13 | 13 | 13 | 13 | 13 | 13 | 13 | 13 | 13 | 13 | 13 | 13 | 13 | 13 | 13 | 13 | 13 | 13 | 13 | 13 | 13 | 13 | 13 | 13 |
| 14 | 14 | 14 | 14 | 14 | 14 | 14 | 14 | 14 | 14 | 14 | 14 | 14 | 14 | 14 | 14 | 14 | 14 | 14 | 14 | 14 | 14 | 14 | 14 | 14 |
| 15 | 15 | 15 | 15 | 15 | 15 | 15 | 15 | 15 | 15 | 15 | 15 | 15 | 15 | 15 | 15 | 15 | 15 | 15 | 15 | 15 | 15 | 15 | 15 | 15 |

*This chart is intended to be used as a guide only. Improper drill use may lead to premature wear and the need to replace the drill prior to 15 uses.

NOTE: Surgical instruments are susceptible to damage for a variety of reasons including but not limited to: prolonged use, misuse, rough or improper handling, inadequate cleaning, etc. Care must be taken to avoid compromising the intended performance of the instrument. Please refer to P-IFSCSS for complete instructions on the sterilization and care of stainless steel.

